

CERTIFICATE

on the antimicrobial properties of materials
The product

ShieldTec Aerosol Lack ST045 AS

of the company

SISTEC Coatings GmbH

MauserstraÙe 6/1
71640 Ludwigsburg

has been successfully tested against
Staphylococcus aureus DSM 345/ATCC 6538P
Escherichia coli DSM 1576/ATCC 8739

as antimicrobial
(>99.99% reduction)

according to ISO 22196 - 2007 (modified)

For details, see Test Reports 2139/ 150708-10208-22196-01; Sample: 102080807150011
2139W.1/ 150819-10208-22196-01; Sample: 102080807150011

The bearer of this certificate is entitled to affix the following seal to
parts or surfaces made from the tested material



Nuremberg, Germany, November 25th, 2015


Managing Director


Laboratory Director

QualityLabs BT GmbH, NeumeyerstraÙe 46a, 90411 Nuremberg, Germany

The issuer of this certificate is an accredited and recognized testing laboratory for testing antimicrobial and adhesive surface properties of materials used in medical devices.

Register number: D-PL-13335-01-00 / ZLG-AP-231.10.72

Competency certification according to directives 93/42/EWG, 90/385/EWG and DIN EN ISO/IEC 17025.